

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
 Data File : VN071333.D
 Acq On : 16 Mar 2022 20:25
 Operator : JC\MD
 Sample : N1957-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 17 05:21:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	856292	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1407647	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1311833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	470175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	535762	45.652	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.300%
35) Dibromofluoromethane	8.022	113	408220	45.957	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.920%
50) Toluene-d8	10.439	98	1715851	47.976	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.960%
62) 4-Bromofluorobenzene	12.727	95	629250	51.958	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.920%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.857	43	14828	1.428	ug/l #	85
30) Chloroform	7.816	83	15136	0.816	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031622\
Data File : VN071333.D
Acq On : 16 Mar 2022 20:25
Operator : JC\MD
Sample : N1957-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 17 05:21:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

