

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031720\
 Data File : VN060548.D
 Acq On : 17 Mar 2020 12:02
 Operator : JC/MD
 Sample : L1946-01 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 40748

Quant Time: Mar 18 07:03:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	191773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	313394	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	277776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	113057	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	119484	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
35) Dibromofluoromethane	7.56	113	90935	62.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.02%	
50) Toluene-d8	10.08	98	381602	51.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.66%	
62) 4-Bromofluorobenzene	12.40	95	131003	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.03	105	10144	1.424	ug/l	99
95) Naphthalene	15.13	128	10460	1.288	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031720\
Data File : VN060548.D
Acq On : 17 Mar 2020 12:02
Operator : JC/MD
Sample : L1946-01 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
40748

Quant Time: Mar 18 07:03:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

