

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031720\
 Data File : VN060550.D
 Acq On : 17 Mar 2020 12:47
 Operator : JC/MD
 Sample : L1943-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8A

Quant Time: Mar 18 07:08:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	335595	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	309674	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136043	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	129983	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
35) Dibromofluoromethane	7.56	113	98469	63.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.42%	
50) Toluene-d8	10.08	98	415580	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
62) 4-Bromofluorobenzene	12.40	95	151323	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
Target Compounds						
16) Acetone	3.79	43	8767	10.460	ug/l	96
20) Methylene Chloride	4.53	84	6891	2.994	ug/l	97
43) Isopropyl Acetate	8.14	43	61901	11.675	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031720\
Data File : VN060550.D
Acq On : 17 Mar 2020 12:47
Operator : JC/MD
Sample : L1943-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8A

Quant Time: Mar 18 07:08:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

