

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031720\
 Data File : VN060551.D
 Acq On : 17 Mar 2020 13:10
 Operator : JC/MD
 Sample : L1943-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9A

Quant Time: Mar 18 07:10:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	344617	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	314682	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	139167	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133543	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
35) Dibromofluoromethane	7.57	113	100538	62.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.70%	
50) Toluene-d8	10.08	98	422078	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
62) 4-Bromofluorobenzene	12.40	95	152268	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
Target Compounds						
16) Acetone	3.79	43	7642	8.993	ug/l	97
20) Methylene Chloride	4.52	84	8433	3.614	ug/l	98
43) Isopropyl Acetate	8.14	43	61010	11.206	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031720\
Data File : VN060551.D
Acq On : 17 Mar 2020 13:10
Operator : JC/MD
Sample : L1943-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9A

Quant Time: Mar 18 07:10:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

