

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031720\
 Data File : VN060557.D
 Acq On : 17 Mar 2020 15:26
 Operator : JC/MD
 Sample : PB127594ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127594ZHE#02

Quant Time: Mar 18 07:30:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	203890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336118	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	312960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	139654	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131976	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
35) Dibromofluoromethane	7.57	113	99647	63.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.74%	
50) Toluene-d8	10.08	98	413886	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
62) 4-Bromofluorobenzene	12.40	95	153291	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
Target Compounds						
16) Acetone	3.79	43	4958	5.869	ug/l	94
20) Methylene Chloride	4.52	84	4571	1.970	ug/l	91
43) Isopropyl Acetate	8.14	43	50939	9.593	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031720\
Data File : VN060557.D
Acq On : 17 Mar 2020 15:26
Operator : JC/MD
Sample : PB127594ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127594ZHE#02

Quant Time: Mar 18 07:30:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

