

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031720\
 Data File : VN060560.D
 Acq On : 17 Mar 2020 16:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 07:41:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.458	0.336	26.6#	70	0.00
3 P	Chloromethane	0.798	0.675	15.4	84	0.00
4 C	Vinyl Chloride	0.684	0.526	23.1#	76	0.00
5 T	Bromomethane	0.301	0.226	24.9	77	0.00
6 T	Chloroethane	0.371	0.330	11.1	89	0.00
7 T	Trichlorofluoromethane	0.875	0.745	14.9	83	0.00
8 T	Diethyl Ether	0.330	0.323	2.1	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.442	10.7	86	0.00
10 T	Methyl Iodide	0.452	0.317	29.9#	68	0.00
11 T	Tert butyl alcohol	0.095	0.101	-6.3	99	0.00
12 CM	1,1-Dichloroethene	0.525	0.436	17.0#	85	0.00
13 T	Acrolein	0.089	0.099	-11.2	99	0.00
14 T	Allyl chloride	0.886	0.810	8.6	87	0.00
15 T	Acrylonitrile	0.251	0.292	-16.3	106	0.00
16 T	Acetone	0.207	0.233	-12.6	108	0.00
17 T	Carbon Disulfide	1.490	1.010	32.2#	74	0.00
18 T	Methyl Acetate	0.691	0.764	-10.6	111	0.00
19 T	Methyl tert-butyl Ether	1.744	1.746	-0.1	97	0.00
20 T	Methylene Chloride	0.569	0.530	6.9	93	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.467	21.5	85	0.00
22 T	Diisopropyl ether	1.794	1.916	-6.8	102	0.00
23 T	Vinyl Acetate	1.426	1.531	-7.4	99	0.00
24 P	1,1-Dichloroethane	1.016	1.011	0.5	96	0.00
25 T	2-Butanone	0.335	0.391	-16.7	109	0.00
26 T	2,2-Dichloropropane	0.925	0.818	11.6	86	0.00
27 T	cis-1,2-Dichloroethene	0.660	0.594	10.0	92	0.00
28 T	Bromochloromethane	0.417	0.427	-2.4	105	0.00
29 T	Tetrahydrofuran	0.233	0.263	-12.9	107	0.00
30 C	Chloroform	1.011	0.991	2.0#	95	0.00
31 T	Cyclohexane	0.942	0.810	14.0	87	0.00
32 T	1,1,1-Trichloroethane	0.896	0.834	6.9	90	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.637	0.8	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.232	0.282	-21.6	123	0.00
36 T	1,1-Dichloropropene	0.477	0.408	14.5	90	0.00
37 T	Ethyl Acetate	0.451	0.485	-7.5	106	0.00
38 T	Carbon Tetrachloride	0.464	0.384	17.2	83	0.00
39 T	Methylcyclohexane	0.562	0.444	21.0	82	0.00
40 TM	Benzene	1.387	1.264	8.9	93	0.00
41 T	Methacrylonitrile	0.208	0.233	-12.0	108	0.00
42 TM	1,2-Dichloroethane	0.500	0.455	9.0	92	0.00
43 T	Isopropyl Acetate	0.790	0.829	-4.9	104	0.00
44 TM	Trichloroethene	0.386	0.319	17.4	86	0.00
45 C	1,2-Dichloropropane	0.363	0.365	-0.6#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031720\
 Data File : VN060560.D
 Acq On : 17 Mar 2020 16:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 18 07:41:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.215	8.5	91	0.00
47 T	Bromodichloromethane	0.479	0.464	3.1	96	0.00
48 T	Methyl methacrylate	0.360	0.364	-1.1	101	0.00
49 T	1,4-Dioxane	0.005	0.004	20.0	91	0.00
50 S	Toluene-d8	1.175	1.102	6.2	98	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.495	-10.5	107	0.00
52 CM	Toluene	0.871	0.780	10.4#	91	0.00
53 T	t-1,3-Dichloropropene	0.586	0.534	8.9	94	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.563	7.4	94	0.00
55 T	1,1,2-Trichloroethane	0.350	0.326	6.9	96	0.00
56 T	Ethyl methacrylate	0.503	0.527	-4.8	100	0.00
57 T	1,3-Dichloropropane	0.571	0.559	2.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.263	0.263	0.0	98	0.00
59 T	2-Hexanone	0.329	0.358	-8.8	105	0.00
60 T	Dibromochloromethane	0.356	0.343	3.7	95	0.00
61 T	1,2-Dibromoethane	0.342	0.320	6.4	93	0.00
62 S	4-Bromofluorobenzene	0.445	0.426	4.3	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.444	0.327	26.4#	80	0.00
65 PM	Chlorobenzene	1.010	0.890	11.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.339	9.1	91	0.00
67 C	Ethyl Benzene	1.839	1.645	10.5#	90	0.00
68 T	m/p-Xylenes	0.694	0.598	13.8	88	0.00
69 T	o-Xylene	0.661	0.592	10.4	91	0.00
70 T	Styrene	1.088	0.990	9.0	90	0.00
71 P	Bromoform	0.269	0.272	-1.1	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.706	3.316	10.5	89	0.00
74 T	N-amyl acetate	1.619	1.635	-1.0	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.007	1.074	-6.7	105	0.00
76 T	1,2,3-Trichloropropane	1.097	0.907	17.3	82	0.00
77 T	Bromobenzene	0.908	0.815	10.2	91	0.00
78 T	n-propylbenzene	4.344	3.881	10.7	90	0.00
79 T	2-Chlorotoluene	2.554	2.306	9.7	92	0.00
80 T	1,3,5-Trimethylbenzene	3.120	2.773	11.1	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.401	-1.0	96	0.00
82 T	4-Chlorotoluene	2.690	2.375	11.7	91	0.00
83 T	tert-Butylbenzene	2.701	2.366	12.4	87	0.00
84 T	1,2,4-Trimethylbenzene	3.149	2.793	11.3	88	0.00
85 T	sec-Butylbenzene	3.618	3.181	12.1	87	0.00
86 T	p-Isopropyltoluene	3.348	2.836	15.3	86	0.00
87 T	1,3-Dichlorobenzene	1.706	1.416	17.0	86	0.00
88 T	1,4-Dichlorobenzene	1.697	1.406	17.1	87	0.00
89 T	n-Butylbenzene	3.074	2.586	15.9	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031720\
Data File : VN060560.D
Acq On : 17 Mar 2020 16:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 18 07:41:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.485	0.431	11.1	86	0.00
91 T	1,2-Dichlorobenzene	1.592	1.387	12.9	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.228	0.220	3.5	97	0.00
93 T	1,2,4-Trichlorobenzene	1.130	0.867	23.3	83	0.00
94 T	Hexachlorobutadiene	0.536	0.408	23.9	84	0.00
95 T	Naphthalene	3.258	2.707	16.9	87	0.00
96 T	1,2,3-Trichlorobenzene	1.062	0.836	21.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6