

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031721\
 Data File : VN066229.D
 Acq On : 17 Mar 2021 12:09
 Operator : JC/MD
 Sample : VN0317MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0317MBL01

Quant Time: Mar 18 04:42:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.582	168	299698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	464202	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	427323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	165904	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	224040	51.89	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.78%	
35) Dibromofluoromethane	7.507	113	157322	53.12	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.24%	
50) Toluene-d8	10.028	98	585285	51.11	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.22%	
62) 4-Bromofluorobenzene	12.345	95	162291	39.32	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 78.64%#	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031721\
Data File : VN066229.D
Acq On : 17 Mar 2021 12:09
Operator : JC/MD
Sample : VN0317MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0317MBL01

Quant Time: Mar 18 04:42:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

