

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031721\
 Data File : VN066239.D
 Acq On : 17 Mar 2021 17:08
 Operator : JC/MD
 Sample : M1687-23
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1048

Quant Time: Mar 18 04:46:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	347251	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.518	114	517359	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	439851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	157643	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.952	65	218492	43.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.34%
35) Dibromofluoromethane	7.509	113	175704	53.24	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.48%
50) Toluene-d8	10.030	98	688351	53.94	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.88%
62) 4-Bromofluorobenzene	12.345	95	208277	45.28	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.56%
Target Compounds						
16) Acetone	3.741	43	11868	4.94	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031721\
Data File : VN066239.D
Acq On : 17 Mar 2021 17:08
Operator : JC/MD
Sample : M1687-23
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1048

Quant Time: Mar 18 04:46:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

