

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031721\
 Data File : VN066240.D
 Acq On : 17 Mar 2021 17:31
 Operator : JC/MD
 Sample : M1687-26
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1056

Quant Time: Mar 18 04:46:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	320227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	482421	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	413269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	161951	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	205943	44.64	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	89.28%		
35) Dibromofluoromethane	7.510	113	165495	53.77	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	107.54%		
50) Toluene-d8	10.028	98	617003	51.85	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	103.70%		
62) 4-Bromofluorobenzene	12.345	95	180579	42.10	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.20%		
Target Compounds						Qvalue
16) Acetone	3.741	43	19147	8.64	ug/l	99
52) Toluene	10.092	92	3699	0.44	ug/l	91
67) Ethyl Benzene	11.455	91	4378	0.30	ug/l	90
68) m/p-Xylenes	11.565	106	5596	1.03	ug/l	93
69) o-Xylene	11.887	106	2339	0.45	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031721\
Data File : VN066240.D
Acq On : 17 Mar 2021 17:31
Operator : JC/MD
Sample : M1687-26
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1056

Quant Time: Mar 18 04:46:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

