

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
 Data File : VN071351.D
 Acq On : 17 Mar 2022 14:43
 Operator : JC\MD
 Sample : N1978-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-TB-20220315

Quant Time: Mar 17 15:23:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 17 06:37:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	113124	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	611785	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.745	117	508927	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.433	65	270469	30.114	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.367%
60) 4-Bromofluorobenzene	12.727	95	189448	24.095	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	80.333%
63) Toluene-d8	10.439	98	730408	31.199	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
Data File : VN071351.D
Acq On : 17 Mar 2022 14:43
Operator : JC\MD
Sample : N1978-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-TB-20220315

Quant Time: Mar 17 15:23:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 17 06:37:08 2022
Response via : Initial Calibration

