

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
 Data File : VN071353.D
 Acq On : 17 Mar 2022 15:58
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 18 07:01:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 17 06:37:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	105970	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	583679	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	489827	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	254855	30.292	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.967%
60) 4-Bromofluorobenzene	12.727	95	186546	24.651	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.167%
63) Toluene-d8	10.439	98	697938	30.974	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.233%
Target Compounds						
					Qvalue	
91) Naphthalene	15.504	128	27102	1.708	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
Data File : VN071353.D
Acq On : 17 Mar 2022 15:58
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 18 07:01:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 17 06:37:08 2022
Response via : Initial Calibration

