

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
 Data File : VN071355.D
 Acq On : 17 Mar 2022 16:46
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 18 07:01:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 17 06:37:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.651	128	110077	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	594693	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	493136	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.433	65	260063	29.757	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.200%
60) 4-Bromofluorobenzene	12.727	95	186605	24.493	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.633%
63) Toluene-d8	10.439	98	698618	30.796	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.667%
Target Compounds						
					Qvalue	
89) 1,2,4-Trichlorobenzene	15.262	180	16149	3.164	ug/l	96
91) Naphthalene	15.504	128	28463	1.782	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
Data File : VN071355.D
Acq On : 17 Mar 2022 16:46
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 18 07:01:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 17 06:37:08 2022
Response via : Initial Calibration

