

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN031722\  
 Data File : VN071360.D  
 Acq On : 17 Mar 2022 18:47  
 Operator : JC\MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 IBLK

Quant Time: Mar 18 07:03:00 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N031722W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Thu Mar 17 06:37:08 2022  
 Response via : Initial Calibration

| Compound                | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------|--------|------|----------|--------|-------|----------|
| -----                   |        |      |          |        |       |          |
| Internal Standards      |        |      |          |        |       |          |
| 1) Bromochloromethane   | 7.657  | 128  | 97858    | 30.000 | ug/l  | 0.00     |
| 28) 1,4-Difluorobenzene | 8.969  | 114  | 559808   | 30.000 | ug/l  | 0.00     |
| 57) Chlorobenzene-d5    | 11.745 | 117  | 479387   | 30.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 27) 1,2-Dichloroethane-d4   | 8.433  | 65    | 247207   | 31.818   | ug/l | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 110 | Recovery | =    | 106.067% |
| 60) 4-Bromofluorobenzene    | 12.727 | 95    | 184537   | 24.917   | ug/l | 0.00     |
| Spiked Amount               | 30.000 | Range | 63 - 112 | Recovery | =    | 83.067%  |
| 63) Toluene-d8              | 10.439 | 98    | 668211   | 30.301   | ug/l | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 112 | Recovery | =    | 101.000% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN031722\  
Data File : VN071360.D  
Acq On : 17 Mar 2022 18:47  
Operator : JC\MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
IBLK

Quant Time: Mar 18 07:03:00 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\624N031722W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Thu Mar 17 06:37:08 2022  
Response via : Initial Calibration

