

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
 Data File : VN071343.D
 Acq On : 17 Mar 2022 02:27
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN031722

Quant Time: Mar 17 06:40:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 17 06:37:08 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	30.000	30.000	0.0	105	0.00
2 M	Dichlorodifluoromethane	20.000	20.073	-0.4	104	0.00
3 M	Chloromethane	20.000	19.606	2.0	102	0.00
4 M	Vinyl Chloride	20.000	19.320	3.4	100	0.00
5 M	Bromomethane	20.000	22.971	-14.9	111	0.00
6 M	Chloroethane	20.000	19.626	1.9	101	0.00
7 M	Trichlorofluoromethane	20.000	19.721	1.4	104	0.00
8 T	Diethyl Ether	20.000	19.307	3.5	105	0.00
9	1,1,2-Trichlorotrifluoroeth	20.000	19.145	4.3	101	0.00
10 M	1,1-Dichloroethene	20.000	19.260	3.7	103	0.00
11	Methyl Iodide	20.000	17.864	10.7	98	0.00
12	Methyl Acetate	20.000	19.252	3.7	104	0.00
13 M	Acrolein	100.000	96.917	3.1	102	0.00
14 M	Acrylonitrile	100.000	94.396	5.6	101	0.00
15 M	Acetone	100.000	96.085	3.9	104	0.00
16 M	Carbon Disulfide	20.000	19.530	2.3	107	0.00
17	Allyl chloride	20.000	18.806	6.0	101	0.00
18 M	Methylene Chloride	20.000	18.971	5.1	101	0.00
19 M	trans-1,2-Dichloroethene	20.000	19.781	1.1	106	0.00
20 T	Diisopropyl ether	20.000	19.533	2.3	107	0.00
21 M	1,1-Dichloroethane	20.000	19.131	4.3	100	0.00
22 M	cis-1,2-Dichloroethene	20.000	19.328	3.4	104	0.00
23 M	tert-Butyl Alcohol	100.000	92.379	7.6	101	0.00
24 M	Methyl tert-Butyl Ether	20.000	19.389	3.1	104	0.00
25 M	Chloroform	20.000	19.361	3.2	101	0.00
26	Cyclohexane	20.000	19.286	3.6	104	0.00
27 s	1,2-Dichloroethane-d4	30.000	29.266	2.4	102	0.00
28 I	1,4-Difluorobenzene	30.000	30.000	0.0	104	0.00
29	1,1-Dichloropropene	20.000	19.564	2.2	104	0.00
30 M	2-Butanone	100.000	95.180	4.8	101	0.00
31	2,2-Dichloropropane	20.000	18.235	8.8	95	0.00
32 M	1,1,1-Trichloroethane	20.000	19.652	1.7	103	0.00
33 M	Carbon Tetrachloride	20.000	19.717	1.4	103	0.00
34 M	Benzene	20.000	19.537	2.3	102	0.00
35	Methacrylonitrile	20.000	20.350	-1.8	106	0.00
36 M	1,2-Dichloroethane	20.000	19.832	0.8	104	0.00
37 M	Trichloroethene	20.000	19.420	2.9	104	0.00
38	Methylcyclohexane	20.000	19.613	1.9	106	0.00
39 M	1,2-Dichloropropane	20.000	19.635	1.8	102	0.00
40	Dibromomethane	20.000	18.950	5.3	99	0.00
41 M	Bromodichloromethane	20.000	19.506	2.5	103	0.00
42 M	Vinyl Acetate	100.000	97.819	2.2	103	0.00
43	Ethyl Acetate	20.000	19.201	4.0	101	0.00
44	Isopropyl Acetate	20.000	18.917	5.4	102	0.00
45 T	1,4-Dioxane	400.000	371.731	7.1	97	0.00
46	Methyl methacrylate	20.000	18.330	8.4	103	0.00
47	n-amyl Acetate	20.000	18.117	9.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
 Data File : VN071343.D
 Acq On : 17 Mar 2022 02:27
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN031722

Quant Time: Mar 17 06:40:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 17 06:37:08 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 M	t-1,3-Dichloropropene	20.000	18.607	7.0	102	0.00
49 T	cis-1,3-Dichloropropene	20.000	19.151	4.2	103	0.00
50 M	1,1,2-Trichloroethane	20.000	19.282	3.6	103	0.00
51	Ethyl methacrylate	20.000	18.751	6.2	103	0.00
52	1,3-Dichloropropane	20.000	19.451	2.7	103	0.00
53 M	Dibromochloromethane	20.000	19.254	3.7	103	0.00
54 M	1,2-Dibromoethane	20.000	19.199	4.0	104	0.00
55 M	2-Chloroethyl vinyl ether	100.000	84.740	15.3	101	0.00
56 M	Bromoform	20.000	18.797	6.0	105	0.00
57 I	Chlorobenzene-d5	30.000	30.000	0.0	103	0.00
58 M	4-Methyl-2-Pentanone	100.000	100.220	-0.2	101	0.00
59 M	2-Hexanone	100.000	98.748	1.3	101	0.00
60 S	4-Bromofluorobenzene	30.000	29.460	1.8	102	0.00
61 M	Tetrachloroethene	20.000	20.171	-0.9	104	0.00
62 M	Toluene	20.000	20.295	-1.5	102	0.00
63 S	Toluene-d8	30.000	30.964	-3.2	104	0.00
64 M	Chlorobenzene	20.000	20.217	-1.1	104	0.00
65	1,1,1,2-Tetrachloroethane	20.000	20.217	-1.1	103	0.00
66 M	Ethyl Benzene	20.000	19.788	1.1	102	0.00
67 M	m/p-Xylenes	40.000	40.246	-0.6	104	0.00
68 M	o-Xylene	20.000	20.050	-0.3	103	0.00
69 M	Styrene	20.000	19.690	1.5	104	0.00
70	Isopropylbenzene	20.000	19.832	0.8	103	0.00
71 M	1,1,2,2-Tetrachloroethane	20.000	20.123	-0.6	101	0.00
72	1,2,3-Trichloropropane	20.000	20.125	-0.6	114	0.00
73	Bromobenzene	20.000	19.658	1.7	104	0.00
74	n-propylbenzene	20.000	19.427	2.9	103	0.00
75	2-Chlorotoluene	20.000	19.543	2.3	102	0.00
76	1,3,5-Trimethylbenzene	20.000	19.892	0.5	103	0.00
77	t-1,4-Dichloro-2-butene	20.000	18.310	8.5	101	0.00
78	4-Chlorotoluene	20.000	19.168	4.2	103	0.00
79	tert-butylbenzene	20.000	19.499	2.5	101	0.00
80	1,2,4-Trimethylbenzene	20.000	19.912	0.4	101	0.00
81	sec-Butylbenzene	20.000	19.547	2.3	105	0.00
82	p-Isopropyltoluene	20.000	19.277	3.6	103	0.00
83 M	1,3-Dichlorobenzene	20.000	19.822	0.9	105	0.00
84 M	1,4-Dichlorobenzene	20.000	19.592	2.0	105	0.00
85	n-Butylbenzene	20.000	18.625	6.9	106	0.00
86 T	Hexachloroethane	20.000	19.123	4.4	105	0.00
87 M	1,2-Dichlorobenzene	20.000	20.051	-0.3	103	0.00
88	1,2-Dibromo-3-Chloropropane	20.000	18.706	6.5	99	0.00
89	1,2,4-Trichlorobenzene	20.000	18.828	5.9	106	0.00
90	Hexachlorobutadiene	20.000	20.383	-1.9	106	0.00
91 M	Naphthalene	20.000	18.634	6.8	112	0.00
92	1,2,3-Trichlorobenzene	20.000	19.082	4.6	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031722\
Data File : VN071343.D
Acq On : 17 Mar 2022 02:27
Operator : JC\MD
Sample : VSTDICV020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN031722

Quant Time: Mar 17 06:40:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N031722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 17 06:37:08 2022
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#)= Out of Range					
SPCC's out = 0 CCC's out = 0					