

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077031.D
 Acq On : 17 Mar 2023 12:14
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 18 02:27:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	360741	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	649655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	544143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	199029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	269812	46.825	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.640%		
35) Dibromofluoromethane	8.172	113	207785	45.671	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.340%		
50) Toluene-d8	10.572	98	775329	47.707	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.420%		
62) 4-Bromofluorobenzene	12.854	95	246322	42.892	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.780%		
Target Compounds						Qvalue
3) Chloromethane	2.372	50	979	0.269	ug/l	94
17) Carbon Disulfide	4.725	76	8826	0.691	ug/l	99
64) Tetrachloroethene	11.101	164	936	0.256	ug/l #	75
65) Chlorobenzene	11.895	112	3658	0.292	ug/l #	77
77) Bromobenzene	12.989	156	1810	0.453	ug/l	98
78) n-propylbenzene	13.042	91	6171	0.300	ug/l	95
82) 4-Chlorotoluene	13.224	91	6193	0.496	ug/l	95
87) 1,3-Dichlorobenzene	13.736	146	4210	0.557	ug/l	86
89) n-Butylbenzene	14.060	91	6779	0.541	ug/l	96
91) 1,2-Dichlorobenzene	14.113	146	4138	0.550	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.724	75	867	0.724	ug/l	75
93) 1,2,4-Trichlorobenzene	15.395	180	5310	1.560	ug/l	97
94) Hexachlorobutadiene	15.501	225	1008	0.538	ug/l	76
95) Naphthalene	15.648	128	23921	3.652	ug/l	97
96) 1,2,3-Trichlorobenzene	15.848	180	6736	1.973	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077031.D
Acq On : 17 Mar 2023 12:14
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 18 02:27:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

