

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077035.D
 Acq On : 17 Mar 2023 15:06
 Operator : JC\MD
 Sample : VN0317MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0317MBL01

Quant Time: Mar 18 02:29:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	385018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	704224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	615662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	233959	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	300044	48.788	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.580%
35) Dibromofluoromethane	8.172	113	227155	46.060	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.120%
50) Toluene-d8	10.572	98	863891	49.038	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.854	95	272157	43.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077035.D
Acq On : 17 Mar 2023 15:06
Operator : JC\MD
Sample : VN0317MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0317MBL01

Quant Time: Mar 18 02:29:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

