

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077043.D
 Acq On : 17 Mar 2023 19:23
 Operator : JC\MD
 Sample : 01941-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB04-20230313

Quant Time: Mar 18 02:33:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	423912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	792998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	677908	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	252238	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	334473	49.396	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.800%
35) Dibromofluoromethane	8.172	113	251439	45.276	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.560%
50) Toluene-d8	10.572	98	950126	47.895	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.780%
62) 4-Bromofluorobenzene	12.854	95	306409	43.711	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077043.D
Acq On : 17 Mar 2023 19:23
Operator : JC\MD
Sample : 01941-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB04-20230313

Quant Time: Mar 18 02:33:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

