

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077045.D
 Acq On : 17 Mar 2023 20:12
 Operator : JC\MD
 Sample : 01904-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE117D1-20230313

Quant Time: Mar 18 02:34:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	403610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	740651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	640257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	236549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	314326	48.756	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.520%		
35) Dibromofluoromethane	8.172	113	232843	44.891	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.780%		
50) Toluene-d8	10.572	98	910696	49.152	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.300%		
62) 4-Bromofluorobenzene	12.854	95	285615	43.624	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.240%		
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.366	96	5203	1.062	ug/l #	60
27) cis-1,2-Dichloroethene	7.501	96	4539	0.746	ug/l	85
30) Chloroform	7.972	83	2855	0.262	ug/l #	82
38) Carbon Tetrachloride	8.366	117	7642	0.968	ug/l	95
44) Trichloroethene	9.360	130	429427	76.091	ug/l	94
55) 1,1,2-Trichloroethane	11.013	97	2259	0.391	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077045.D
Acq On : 17 Mar 2023 20:12
Operator : JC\MD
Sample : 01904-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE117D1-20230313

Quant Time: Mar 18 02:34:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

