

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077046.D
 Acq On : 17 Mar 2023 20:36
 Operator : JC\MD
 Sample : 01904-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE117D2-20230313

Quant Time: Mar 18 02:35:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	341378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	629284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	548242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	202743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	276592	50.724	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.440%
35) Dibromofluoromethane	8.172	113	204504	46.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.800%
50) Toluene-d8	10.571	98	767906	48.780	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.560%
62) 4-Bromofluorobenzene	12.854	95	244613	43.973	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.940%
Target Compounds						
						Qvalue
44) Trichloroethene	9.360	130	5119	1.068	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077046.D
Acq On : 17 Mar 2023 20:36
Operator : JC\MD
Sample : 01904-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE117D2-20230313

Quant Time: Mar 18 02:35:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

