

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077051.D
 Acq On : 17 Mar 2023 22:38
 Operator : JC\MD
 Sample : 01941-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT172S1-20230313

Quant Time: Mar 18 02:37:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	348758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	636505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	576252	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	221689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	278493	49.992	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.980%
35) Dibromofluoromethane	8.172	113	211998	47.560	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	10.571	98	794979	49.927	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.860%
62) 4-Bromofluorobenzene	12.854	95	259891	46.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.380%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.384	101	4700	1.015	ug/l #	68
30) Chloroform	7.978	83	3539	0.375	ug/l	97
44) Trichloroethene	9.360	130	144530	29.800	ug/l	100
64) Tetrachloroethene	11.107	164	5206	1.346	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077051.D
Acq On : 17 Mar 2023 22:38
Operator : JC\MD
Sample : 01941-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT172S1-20230313

Quant Time: Mar 18 02:37:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

