

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077053.D
 Acq On : 17 Mar 2023 23:26
 Operator : JC\MD
 Sample : 01941-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT161S1-20230314

Quant Time: Mar 18 02:37:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	324788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	592046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	516621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	188147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	266530	51.375	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.760%
35) Dibromofluoromethane	8.178	113	195028	47.038	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.080%
50) Toluene-d8	10.572	98	727449	49.117	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.240%
62) 4-Bromofluorobenzene	12.854	95	227235	43.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.840%
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	5.807	73	5426	0.383	ug/l	94
24) 1,1-Dichloroethane	6.578	63	24640	2.722	ug/l	96
30) Chloroform	7.978	83	2543	0.289	ug/l	87
44) Trichloroethene	9.360	130	3979	0.882	ug/l	86
64) Tetrachloroethene	11.107	164	1111	0.320	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077053.D
Acq On : 17 Mar 2023 23:26
Operator : JC\MD
Sample : 01941-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT161S1-20230314

Quant Time: Mar 18 02:37:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

