

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077055.D
 Acq On : 18 Mar 2023 00:15
 Operator : JC\MD
 Sample : 01941-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT162S1-20230314

Quant Time: Mar 18 02:38:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.237	168	331886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	613816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	516805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	189382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	270712	51.065	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.140%
35) Dibromofluoromethane	8.172	113	204267	47.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.040%
50) Toluene-d8	10.572	98	742459	48.352	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.700%
62) 4-Bromofluorobenzene	12.854	95	225691	41.594	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.180%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.360	96	2227	0.553	ug/l	86
27) cis-1,2-Dichloroethene	7.501	96	1357	0.271	ug/l	68
30) Chloroform	7.972	83	3636	0.405	ug/l	88
44) Trichloroethene	9.360	130	61645	13.180	ug/l	91
64) Tetrachloroethene	11.113	164	3907	1.126	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077055.D
Acq On : 18 Mar 2023 00:15
Operator : JC\MD
Sample : 01941-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT162S1-20230314

Quant Time: Mar 18 02:38:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

