

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
 Data File : VN077056.D
 Acq On : 18 Mar 2023 00:39
 Operator : JC\MD
 Sample : 01941-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE114D2-20230314

Quant Time: Mar 18 02:39:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 18 02:17:37 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	440860	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	812146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	707939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	251406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	352745	50.092	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.180%		
35) Dibromofluoromethane	8.178	113	263378	46.308	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.620%		
50) Toluene-d8	10.572	98	1001996	49.319	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.640%		
62) 4-Bromofluorobenzene	12.854	95	313600	43.682	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 87.360%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.378	101	64657	11.051	ug/l	99
12) 1,1-Dichloroethene	4.366	96	4027	0.752	ug/l #	85
27) cis-1,2-Dichloroethene	7.501	96	8511	1.281	ug/l	89
30) Chloroform	7.966	83	3901	0.327	ug/l #	70
38) Carbon Tetrachloride	8.366	117	3432	0.397	ug/l #	83
44) Trichloroethene	9.360	130	357667	57.797	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031723\
Data File : VN077056.D
Acq On : 18 Mar 2023 00:39
Operator : JC\MD
Sample : 01941-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE114D2-20230314

Quant Time: Mar 18 02:39:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031723W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 18 02:17:37 2023
Response via : Initial Calibration

