

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
 Data File : VN054386.D
 Acq On : 18 Mar 2019 11:42
 Operator : JC/SP
 Sample : K1960-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-MW-4-9.0-031519

Quant Time: Mar 19 05:52:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1919207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2705515	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2237750	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	898853	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	920513	45.16	ug/l	0.00
Spiked Amount			Recovery	=	90.32%	
35) Dibromofluoromethane	7.59	113	816370	46.78	ug/l	0.00
Spiked Amount			Recovery	=	93.56%	
50) Toluene-d8	10.09	98	3079083	47.78	ug/l	0.00
Spiked Amount			Recovery	=	95.56%	
62) 4-Bromofluorobenzene	12.40	95	938178	41.61	ug/l	0.00
Spiked Amount			Recovery	=	83.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	1555531	266.548	ug/l	99
19) Methyl tert-butyl Ether	5.05	73	137808	2.788	ug/l	93
31) Cyclohexane	7.65	56	106592	2.307	ug/l #	78
39) Methylcyclohexane	9.08	83	96824	3.158	ug/l	99
67) Ethyl Benzene	11.51	91	26482	0.325	ug/l	99
68) m/p-Xylenes	11.62	106	48179	1.513	ug/l	91
69) o-Xylene	11.95	106	21171	0.682	ug/l	96
73) Isopropylbenzene	12.25	105	653648	9.089	ug/l	100
78) n-propylbenzene	12.59	91	675235	8.653	ug/l	99
83) tert-Butylbenzene	12.99	119	101743	1.848	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	84191	1.435	ug/l	100
85) sec-Butylbenzene	13.17	105	457792	6.411	ug/l	82
95) Naphthalene	15.13	128	57485	4.827	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
Data File : VN054386.D
Acq On : 18 Mar 2019 11:42
Operator : JC/SP
Sample : K1960-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-MW-4-9.0-031519

Quant Time: Mar 19 05:52:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

