

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
 Data File : VN054387.D
 Acq On : 18 Mar 2019 12:10
 Operator : JC/SP
 Sample : K1960-11DL 40X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-MW-4-9.0-031519DL

Quant Time: Mar 19 05:55:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1906355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2703453	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2233322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	800161	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	903243	44.61	ug/l	0.00
Spiked Amount			Recovery	=	89.22%	
35) Dibromofluoromethane	7.59	113	823593	47.23	ug/l	0.00
Spiked Amount			Recovery	=	94.46%	
50) Toluene-d8	10.09	98	3109602	48.29	ug/l	0.00
Spiked Amount			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	12.40	95	906040	40.21	ug/l	0.00
Spiked Amount			Recovery	=	80.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	81398	14.042	ug/l	99
68) m/p-Xylenes	11.62	106	19540	0.615	ug/l	99
73) Isopropylbenzene	12.25	105	17831	0.279	ug/l	97
78) n-propylbenzene	12.59	91	19341	0.278	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	31903	0.611	ug/l	98
95) Naphthalene	15.14	128	20166	4.004	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031819\
Data File : VN054387.D
Acq On : 18 Mar 2019 12:10
Operator : JC/SP
Sample : K1960-11DL 40X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-MW-4-9.0-031519DL

Quant Time: Mar 19 05:55:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

