

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
 Data File : VN054393.D
 Acq On : 18 Mar 2019 14:54
 Operator : JC/SP
 Sample : K1960-04DL
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T11-MW-9-9.0-031419DL

Quant Time: Mar 19 09:57:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1919392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2739007	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2309606	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	829089	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	921940	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
35) Dibromofluoromethane	7.59	113	821954	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	10.09	98	3151781	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
62) 4-Bromofluorobenzene	12.40	95	943170	41.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.64%	
Target Compounds						
16) Acetone	3.82	43	1939849	332.371	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031819\
Data File : VN054393.D
Acq On : 18 Mar 2019 14:54
Operator : JC/SP
Sample : K1960-04DL
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T11-MW-9-9.0-031419DL

Quant Time: Mar 19 09:57:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

