

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
 Data File : VN054400.D
 Acq On : 18 Mar 2019 18:06
 Operator : JC/SP
 Sample : K1941-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-2

Quant Time: Mar 19 06:52:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1919195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2712046	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2230937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	764814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	909056	44.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.20%	
35) Dibromofluoromethane	7.59	113	825004	47.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.34%	
50) Toluene-d8	10.09	98	3119503	48.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.58%	
62) 4-Bromofluorobenzene	12.40	95	894338	39.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
Data File : VN054400.D
Acq On : 18 Mar 2019 18:06
Operator : JC/SP
Sample : K1941-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-2

Quant Time: Mar 19 06:52:16 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

