

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
 Data File : VN054403.D
 Acq On : 18 Mar 2019 19:28
 Operator : JC/SP
 Sample : K1972-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-8-031519

Quant Time: Mar 19 06:57:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1818364	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2574657	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2130476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	753209	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	867398	44.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.84%	
35) Dibromofluoromethane	7.59	113	789274	47.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.06%	
50) Toluene-d8	10.09	98	2960665	48.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.56%	
62) 4-Bromofluorobenzene	12.40	95	856920	39.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.86%	
Target Compounds						
16) Acetone	3.83	43	15430	2.791	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
Data File : VN054403.D
Acq On : 18 Mar 2019 19:28
Operator : JC/SP
Sample : K1972-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-8-031519

Quant Time: Mar 19 06:57:59 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

