

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031819\
 Data File : VN054407.D
 Acq On : 18 Mar 2019 22:11
 Operator : JC/SP
 Sample : VN0318WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0318WBL02

Quant Time: Mar 19 07:06:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1797930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2589226	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2146882	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	781993	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	867809	45.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.90%	
35) Dibromofluoromethane	7.59	113	785556	47.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.08%	
50) Toluene-d8	10.09	98	2983700	48.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.76%	
62) 4-Bromofluorobenzene	12.40	95	876236	40.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031819\
Data File : VN054407.D
Acq On : 18 Mar 2019 22:11
Operator : JC/SP
Sample : VN0318WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0318WBL02

Quant Time: Mar 19 07:06:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

