

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031819\
 Data File : VN054428.D
 Acq On : 19 Mar 2019 7:46
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 08:15:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.474	0.436	8.0	71	0.00
3 P	Chloromethane	0.460	0.433	5.9	79	0.00
4 C	Vinyl Chloride	0.526	0.457	13.1#	72	0.00
5 T	Bromomethane	0.364	0.293	19.5	71	0.00
6 T	Chloroethane	0.318	0.270	15.1	73	0.00
7 T	Trichlorofluoromethane	0.699	0.575	17.7	71	0.00
8 T	Diethyl Ether	0.319	0.316	0.9	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.524	0.465	11.3	69	0.00
10 T	Methyl Iodide	0.690	0.734	-6.4	80	0.00
11 T	Tert butyl alcohol	0.036	0.036	0.0	82	0.00
12 CM	1,1-Dichloroethene	0.519	0.480	7.5#	73	0.00
13 T	Acrolein	0.034	0.030	11.8	67	0.00
14 T	Allyl chloride	0.666	0.660	0.9	77	0.00
15 T	Acrylonitrile	0.150	0.182	-21.3	95	0.00
16 T	Acetone	0.152	0.108	28.9#	57	0.00
17 T	Carbon Disulfide	1.302	1.089	16.4	64	0.00
18 T	Methyl Acetate	0.308	0.415	-34.7#	102	0.00
19 T	Methyl tert-butyl Ether	1.288	1.461	-13.4	89	0.00
20 T	Methylene Chloride	0.499	0.552	-10.6	90	0.00
21 T	trans-1,2-Dichloroethene	0.498	0.517	-3.8	83	0.00
22 T	Diisopropyl ether	1.330	1.476	-11.0	87	0.00
23 T	Vinyl Acetate	0.981	1.023	-4.3	79	0.00
24 P	1,1-Dichloroethane	0.854	0.938	-9.8	87	0.00
25 T	2-Butanone	0.178	0.199	-11.8	84	0.00
26 T	2,2-Dichloropropane	0.751	0.435	42.1#	46#	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.613	-10.5	86	0.00
28 T	Bromochloromethane	0.349	0.377	-8.0	86	0.00
29 T	Tetrahydrofuran	0.115	0.131	-13.9	90	0.00
30 C	Chloroform	0.891	0.960	-7.7#	85	0.00
31 T	Cyclohexane	0.919	0.797	13.3	80	0.00
32 T	1,1,1-Trichloroethane	0.805	0.804	0.1	77	0.00
33 S	1,2-Dichloroethane-d4	0.531	0.521	1.9	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.322	0.326	-1.2	86	0.00
36 T	1,1-Dichloropropene	0.460	0.450	2.2	81	0.00
37 T	Ethyl Acetate	0.265	0.279	-5.3	87	0.00
38 T	Carbon Tetrachloride	0.483	0.446	7.7	76	0.00
39 T	Methylcyclohexane	0.567	0.504	11.1	73	0.00
40 TM	Benzene	1.318	1.393	-5.7	87	0.00
41 T	Methacrylonitrile	0.151	0.156	-3.3	87	0.00
42 TM	1,2-Dichloroethane	0.445	0.423	4.9	79	0.00
43 T	Isopropyl Acetate	0.464	0.475	-2.4	83	0.00
44 TM	Trichloroethene	0.409	0.409	0.0	83	0.00
45 C	1,2-Dichloropropane	0.327	0.353	-8.0#	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031819\
 Data File : VN054428.D
 Acq On : 19 Mar 2019 7:46
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 08:15:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.221	0.225	-1.8	85	0.00
47 T	Bromodichloromethane	0.443	0.446	-0.7	81	0.00
48 T	Methyl methacrylate	0.236	0.247	-4.7	82	0.00
49 T	1,4-Dioxane	0.003	0.004	-33.3#	96	0.00
50 S	Toluene-d8	1.191	1.185	0.5	85	0.00
51 T	4-Methyl-2-Pentanone	0.246	0.273	-11.0	89	0.00
52 CM	Toluene	0.835	0.871	-4.3#	87	0.00
53 T	t-1,3-Dichloropropene	0.451	0.409	9.3	72	0.00
54 T	cis-1,3-Dichloropropene	0.516	0.497	3.7	77	0.00
55 T	1,1,2-Trichloroethane	0.305	0.329	-7.9	89	0.00
56 T	Ethyl methacrylate	0.364	0.416	-14.3	93	0.00
57 T	1,3-Dichloropropane	0.499	0.537	-7.6	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.138	0.167	-21.0	99	0.00
59 T	2-Hexanone	0.164	0.174	-6.1	87	0.00
60 T	Dibromochloromethane	0.343	0.356	-3.8	82	0.00
61 T	1,2-Dibromoethane	0.310	0.336	-8.4	88	0.00
62 S	4-Bromofluorobenzene	0.417	0.412	1.2	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.483	0.473	2.1	82	0.00
65 PM	Chlorobenzene	1.098	1.154	-5.1	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.399	0.420	-5.3	83	0.00
67 C	Ethyl Benzene	1.820	1.864	-2.4#	84	0.00
68 T	m/p-Xylenes	0.711	0.740	-4.1	85	0.00
69 T	o-Xylene	0.693	0.743	-7.2	87	0.00
70 T	Styrene	1.087	1.176	-8.2	86	0.00
71 P	Bromoform	0.264	0.274	-3.8	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	4.000	3.771	5.7	84	0.00
74 T	N-amyl acetate	0.888	0.904	-1.8	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.852	0.879	-3.2	91	0.00
76 T	1,2,3-Trichloropropane	0.735	0.747	-1.6	89	0.00
77 T	Bromobenzene	1.059	2.705	-155.4#	228#	0.00
78 T	n-propylbenzene	4.341	4.027	7.2	81	0.00
79 T	2-Chlorotoluene	2.610	2.431	6.9	83	0.00
80 T	1,3,5-Trimethylbenzene	3.327	3.191	4.1	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.157	27.0#	61	0.00
82 T	4-Chlorotoluene	2.606	2.475	5.0	84	0.00
83 T	tert-Butylbenzene	3.063	2.816	8.1	83	0.00
84 T	1,2,4-Trimethylbenzene	3.264	3.169	2.9	85	0.00
85 T	sec-Butylbenzene	3.972	3.560	10.4	80	0.00
86 T	p-Isopropyltoluene	3.485	3.171	9.0	80	0.00
87 T	1,3-Dichlorobenzene	1.784	1.718	3.7	84	0.00
88 T	1,4-Dichlorobenzene	1.752	1.689	3.6	84	0.00
89 T	n-Butylbenzene	2.648	2.402	9.3	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN031819\
Data File : VN054428.D
Acq On : 19 Mar 2019 7:46
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 19 08:15:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.577	0.468	18.9	69	0.00
91 T	1,2-Dichlorobenzene	1.744	1.741	0.2	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.132	0.134	-1.5	88	0.00
93 T	1,2,4-Trichlorobenzene	0.901	1.006	-11.7	89	0.00
94 T	Hexachlorobutadiene	0.664	0.512	22.9	72	0.00
95 T	Naphthalene	1.854	2.407	-29.8#	105	0.00
96 T	1,2,3-Trichlorobenzene	0.926	0.994	-7.3	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6