

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031820\
 Data File : VN060590.D
 Acq On : 18 Mar 2020 14:32
 Operator : JC/MD
 Sample : L1917-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-5

Quant Time: Mar 19 05:55:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195920	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	322069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	293977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	127555	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124692	46.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.72%	
35) Dibromofluoromethane	7.57	113	95395	48.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.96%	
50) Toluene-d8	10.08	98	392815	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
62) 4-Bromofluorobenzene	12.40	95	141236	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031820\
Data File : VN060590.D
Acq On : 18 Mar 2020 14:32
Operator : JC/MD
Sample : L1917-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-5

Quant Time: Mar 19 05:55:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

