

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031820\
 Data File : VN060591.D
 Acq On : 18 Mar 2020 14:55
 Operator : JC/MD
 Sample : L1917-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-725

Quant Time: Mar 19 05:58:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	317763	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	295212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	129930	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124353	46.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.84%	
35) Dibromofluoromethane	7.57	113	94769	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	10.08	98	392367	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	12.40	95	142300	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
Target Compounds						
20) Methylene Chloride	4.52	84	2226	0.954	ug/l	89
52) Toluene	10.14	92	2401	0.428	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031820\
Data File : VN060591.D
Acq On : 18 Mar 2020 14:55
Operator : JC/MD
Sample : L1917-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-725

Quant Time: Mar 19 05:58:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

