

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031820\
 Data File : VN060616.D
 Acq On : 19 Mar 2020 1:28
 Operator : JC/MD
 Sample : PB127624ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127624ZHE#10

Quant Time: Mar 19 08:56:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	210076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	352803	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	328460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	146980	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133884	46.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.86%	
35) Dibromofluoromethane	7.56	113	103938	48.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.44%	
50) Toluene-d8	10.08	98	435658	50.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.40	95	161724	50.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.28%	
Target Compounds						
16) Acetone	3.79	43	6009	5.882	ug/l	89
20) Methylene Chloride	4.52	84	6172	2.456	ug/l	91
43) Isopropyl Acetate	8.14	43	61270	10.065	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031820\
Data File : VN060616.D
Acq On : 19 Mar 2020 1:28
Operator : JC/MD
Sample : PB127624ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127624ZHE#10

Quant Time: Mar 19 08:56:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

