

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031820\
 Data File : VN060619.D
 Acq On : 19 Mar 2020 2:36
 Operator : JC/MD
 Sample : L1954-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10A

Quant Time: Mar 19 09:03:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	334733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	312015	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127909	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
35) Dibromofluoromethane	7.56	113	97858	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
50) Toluene-d8	10.08	98	408670	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
62) 4-Bromofluorobenzene	12.40	95	153078	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
Target Compounds						
16) Acetone	3.79	43	7372	7.574	ug/l	92
20) Methylene Chloride	4.52	84	7623	3.184	ug/l	91
43) Isopropyl Acetate	8.14	43	63028	10.912	ug/l #	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031820\
Data File : VN060619.D
Acq On : 19 Mar 2020 2:36
Operator : JC/MD
Sample : L1954-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10A

Quant Time: Mar 19 09:03:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. -0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36

Instrument :

MSVOA_N

ClientSampleId :


TP-10A

Tot Ion:168 Resp: 200177

Ion Ratio Lower Upper

168 100

99 56.3 44.0 66.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.64

80000

60000



40000

20000

0

7.60 7.70

Time-->



#16

Acetone

Concen: 7.574 ug/l

RT: 3.79 min Scan# 626

Delta R.T. 0.01 min

Lab File: VN060619.D


Acq: 19 Mar 2020 2:36

Tgt Ion: 43 Resp: 7372

Ion Ratio Lower Upper

43 100

58 27.1 25.1 37.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

3.79

2500

2000

1500


1000




500

0

3.70 3.80 3.90

Time-->









#20
Methylene Chloride
Concen: 3.184 ug/l
RT: 4.52 min Scan# 855
Delta R.T. 0.00 min
Lab File: VN060619.D
Acq: 19 Mar 2020 2:36


Instrument :
MSVOA_N
ClientSampleId :
TP-10A


Tot Ion:	84	Resp:	7623
Ion	Ratio	Lower	Upper
84	100		
49	114.3	103.3	154.9
51	37.3	31.6	47.4
86	62.0	52.0	78.0



#33
1,2-Dichloroethane-d4
Concen: 46.547 ug/l
RT: 8.00 min Scan# 1937
Delta R.T. -0.00 min
Lab File: VN060619.D
Acq: 19 Mar 2020 2:36

Tgt Ion:	65	Resp:	127909
Ion	Ratio	Lower	Upper
65	100		
67	53.7	0.0	107.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2113

Delta R.T. 0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36

Instrument :

MSVOA_N

ClientSampleId :

TP-10A


Tot Ion:114 Resp: 334733

Ion Ratio Lower Upper

114 100

63 21.9 0.0 43.8




88 16.6 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 48.348 ug/l

RT: 7.56 min Scan# 1801

Delta R.T. 0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36


Tgt Ion:113 Resp: 97858

Ion Ratio Lower Upper

113 100

111 101.9 82.9 124.3



192 19.8 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 10.912 ug/l

RT: 8.14 min Scan# 1980

Delta R.T. 0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36

Instrument :

MSVOA_N

ClientSampleId :

TP-10A





Tot Ion: 43 Resp: 63028

Ion Ratio Lower Upper

43 100

61 18.7 20.7 31.1#

87 12.4 9.8 14.8



#50

Toluene-d8

Concen: 49.432 ug/l

RT: 10.08 min Scan# 2582

Delta R.T. -0.00 min

Lab File: VN060619.D




Acq: 19 Mar 2020 2:36


Tgt Ion: 98 Resp: 408670

Ion Ratio Lower Upper

98 100

100 65.3 51.8 77.8





#62

4-Bromofluorobenzene

Concen: 50.523 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36

Instrument :

MSVOA_N

ClientSampleId :

TP-10A


Tot Ion: 95 Resp: 153078

Ion Ratio Lower Upper

95 100

174 78.2 0.0 157.6




176 75.6 0.0 150.6



Abundance Ion 94.90 (94.60 to 95.60): VN060619.D (-3290) (-)

Ion 173.80 (173.50 to 174.50): VN060619.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN060619.D (-3290) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2994

Delta R.T. 0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36


Tgt Ion: 117 Resp: 312015

Ion Ratio Lower Upper

117 100

82 57.8 46.7 70.1



119 32.4 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): VN060619.D (-2979) (-)

Ion 82.00 (81.70 to 82.70): VN060619.D (-2979) (-)

Ion 118.90 (118.60 to 119.60): VN060619.D (-2979) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN060619.D

Acq: 19 Mar 2020 2:36

Instrument :

MSVOA_N

ClientSampleId :

TP-10A


Tot Ion:152 Resp: 142803

Ion Ratio Lower Upper

152 100

115 60.5 30.2 90.6

150 156.4 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

