

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066249.D
 Acq On : 18 Mar 2021 12:51
 Operator : JC/MD
 Sample : M1714-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-031721

Quant Time: Mar 19 04:23:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	316823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	487111	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	410414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	158179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	215588	47.23	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.46%
35) Dibromofluoromethane	7.507	113	169215	54.45	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.90%
50) Toluene-d8	10.028	98	633758	52.74	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.48%
62) 4-Bromofluorobenzene	12.345	95	190472	43.98	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.96%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066249.D
Acq On : 18 Mar 2021 12:51
Operator : JC/MD
Sample : M1714-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-031721

Quant Time: Mar 19 04:23:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

