

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066250.D
 Acq On : 18 Mar 2021 13:14
 Operator : JC/MD
 Sample : M1714-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR3-224-234-031721

Quant Time: Mar 19 04:24:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	315136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	479390	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	352740	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	138232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	207782	45.76	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.52%
35) Dibromofluoromethane	7.509	113	165708	54.18	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.36%
50) Toluene-d8	10.028	98	538999	45.58	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.16%#
62) 4-Bromofluorobenzene	12.345	95	164890	38.68	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.36%#
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	3.663	96	109985	37.17	ug/l	90
16) Acetone	3.749	43	8050	3.69	ug/l	97
24) 1,1-Dichloroethane	5.747	63	18153	2.53	ug/l	97
27) cis-1,2-Dichloroethene	6.734	96	71082	18.81	ug/l	77
30) Chloroform	7.287	83	22404	3.42	ug/l	94
42) 1,2-Dichloroethane	8.051	62	9883	1.87	ug/l	82
44) Trichloroethene	8.767	130	55521	15.67	ug/l	94
49) 1,4-Dioxane	9.137	88	1949	35.32	ug/l #	77
52) Toluene	10.095	92	112181	13.28	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066250.D
Acq On : 18 Mar 2021 13:14
Operator : JC/MD
Sample : M1714-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR3-224-234-031721

Quant Time: Mar 19 04:24:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

