

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066254.D
 Acq On : 18 Mar 2021 14:44
 Operator : JC/MD
 Sample : PB135215ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135215ZHE#01

Quant Time: Mar 19 04:25:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	303464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	456984	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	430479	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	161213	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	198111	45.31	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.62%		
35) Dibromofluoromethane	7.507	113	162776	55.83	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	111.66%		
50) Toluene-d8	10.028	98	620185	55.02	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	110.04%		
62) 4-Bromofluorobenzene	12.345	95	178829	44.01	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.02%		
Target Compounds					Qvalue	
16) Acetone	3.741	43	85620	40.76	ug/l	99
20) Methylene Chloride	4.457	84	16852	5.13	ug/l	90
43) Isopropyl Acetate	8.094	43	30901	3.02	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066254.D
Acq On : 18 Mar 2021 14:44
Operator : JC/MD
Sample : PB135215ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135215ZHE#01

Quant Time: Mar 19 04:25:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

