

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066255.D
 Acq On : 18 Mar 2021 15:07
 Operator : JC/MD
 Sample : PB135215ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135215ZHE#02

Quant Time: Mar 19 04:26:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	301726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	461315	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	416094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	156682	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	215659	49.61	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.22%		
35) Dibromofluoromethane	7.509	113	168195	57.15	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	114.30%		
50) Toluene-d8	10.028	98	647505	56.90	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	113.80%#		
62) 4-Bromofluorobenzene	12.345	95	180520	44.01	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.02%		
Target Compounds						Qvalue
16) Acetone	3.741	43	80657	38.62	ug/l	97
20) Methylene Chloride	4.455	84	13870	4.25	ug/l	89
43) Isopropyl Acetate	8.091	43	31181	3.02	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066255.D
Acq On : 18 Mar 2021 15:07
Operator : JC/MD
Sample : PB135215ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135215ZHE#02

Quant Time: Mar 19 04:26:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

