

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066259.D
 Acq On : 18 Mar 2021 16:32
 Operator : JC/MD
 Sample : PB135215ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135215ZHE#06

Quant Time: Mar 19 04:27:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	290560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.518	114	431977	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	406780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	145771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.952	65	197590	47.20	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.40%		
35) Dibromofluoromethane	7.512	113	158071	57.36	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	114.72%		
50) Toluene-d8	10.030	98	601395	56.44	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	112.88%#		
62) 4-Bromofluorobenzene	12.345	95	161996	42.18	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.36%		
Target Compounds						
16) Acetone	3.747	43	52761	26.24	ug/l	99
20) Methylene Chloride	4.455	84	13087	4.16	ug/l	96
43) Isopropyl Acetate	8.097	43	28210	2.92	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066259.D
Acq On : 18 Mar 2021 16:32
Operator : JC/MD
Sample : PB135215ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135215ZHE#06

Quant Time: Mar 19 04:27:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

