

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066261.D
 Acq On : 18 Mar 2021 17:18
 Operator : JC/MD
 Sample : PB135215ZHE#08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135215ZHE#08

Quant Time: Mar 19 04:28:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	311780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	467752	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	452482	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	159881	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	213700	47.57	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.14%		
35) Dibromofluoromethane	7.509	113	170331	57.08	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	114.16%		
50) Toluene-d8	10.028	98	671769	58.22	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	116.44%#		
62) 4-Bromofluorobenzene	12.345	95	183460	44.11	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.22%		
Target Compounds						Qvalue
16) Acetone	3.741	43	85108	39.44	ug/l	100
20) Methylene Chloride	4.457	84	17323	5.13	ug/l	90
43) Isopropyl Acetate	8.094	43	30788	2.94	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066261.D
Acq On : 18 Mar 2021 17:18
Operator : JC/MD
Sample : PB135215ZHE#08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135215ZHE#08

Quant Time: Mar 19 04:28:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

