

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066264.D
 Acq On : 18 Mar 2021 18:28
 Operator : JC/MD
 Sample : PB135215ZHE#11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135215ZHE#11

Quant Time: Mar 19 04:29:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	309094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	465146	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	456153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	160047	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	212687	47.76	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.52%		
35) Dibromofluoromethane	7.509	113	171590	57.82	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	115.64%		
50) Toluene-d8	10.028	98	659227	57.46	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	114.92%#		
62) 4-Bromofluorobenzene	12.345	95	183885	44.46	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.92%		
Target Compounds					Qvalue	
16) Acetone	3.744	43	82060	38.36	ug/l	96
20) Methylene Chloride	4.460	84	13934	4.16	ug/l #	91
43) Isopropyl Acetate	8.094	43	30399	2.92	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066264.D
Acq On : 18 Mar 2021 18:28
Operator : JC/MD
Sample : PB135215ZHE#11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135215ZHE#11

Quant Time: Mar 19 04:29:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

