

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066265.D
 Acq On : 18 Mar 2021 18:51
 Operator : JC/MD
 Sample : PB135215TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135215TB

Quant Time: Mar 19 04:29:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	302336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	457100	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	435698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	155594	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	209825	48.17	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.34%
35) Dibromofluoromethane	7.507	113	169776	58.22	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.44%
50) Toluene-d8	10.028	98	644352	57.15	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	114.30%#
62) 4-Bromofluorobenzene	12.345	95	178199	43.85	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.70%
Target Compounds						
16) Acetone	3.741	43	55667	26.60	ug/l	98
20) Methylene Chloride	4.455	84	14779	4.51	ug/l	85
43) Isopropyl Acetate	8.094	43	29161	2.85	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066265.D
Acq On : 18 Mar 2021 18:51
Operator : JC/MD
Sample : PB135215TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135215TB

Quant Time: Mar 19 04:29:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

