

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066266.D
 Acq On : 18 Mar 2021 19:15
 Operator : JC/MD
 Sample : M1693-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5A

Quant Time: Mar 19 04:30:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	285518	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	445365	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	446205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	153885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	232594	56.54	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.08%	
35) Dibromofluoromethane	7.509	113	166450	58.58	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.16%	
50) Toluene-d8	10.028	98	647684	58.96	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 117.92%#	
62) 4-Bromofluorobenzene	12.345	95	183875	46.43	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.86%	
Target Compounds						
16) Acetone	3.741	43	77984	39.46	ug/l	98
20) Methylene Chloride	4.455	84	16414	5.31	ug/l	95
43) Isopropyl Acetate	8.094	43	30701	3.08	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066266.D
Acq On : 18 Mar 2021 19:15
Operator : JC/MD
Sample : M1693-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5A

Quant Time: Mar 19 04:30:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

