

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066267.D
 Acq On : 18 Mar 2021 19:38
 Operator : JC/MD
 Sample : M1687-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1052-1053

Quant Time: Mar 19 04:30:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	313623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	466235	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	404870	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	198743	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	217692	48.18	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.36%		
35) Dibromofluoromethane	7.507	113	172741	58.08	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	116.16%		
50) Toluene-d8	10.028	98	608825	52.94	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	105.88%		
62) 4-Bromofluorobenzene	12.345	95	215557	52.00	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.00%		
Target Compounds					Qvalue	
16) Acetone	3.741	43	95231	43.87	ug/l	96
20) Methylene Chloride	4.455	84	18150	5.34	ug/l	88
39) Methylcyclohexane	9.014	83	11753	2.32	ug/l	95
43) Isopropyl Acetate	8.091	43	27622	2.65	ug/l	95
52) Toluene	10.092	92	247706	30.15	ug/l	100
67) Ethyl Benzene	11.454	91	417211	28.96	ug/l	99
68) m/p-Xylenes	11.559	106	544974	101.95	ug/l	98
69) o-Xylene	11.892	106	388163	75.59	ug/l	96
73) Isopropylbenzene	12.192	105	100605	7.22	ug/l	97
78) n-propylbenzene	12.535	91	218891	13.66	ug/l	97
80) 1,3,5-Trimethylbenzene	12.677	105	314437	27.21	ug/l	95
84) 1,2,4-Trimethylbenzene	12.983	105	1260253	111.56	ug/l	94
85) sec-Butylbenzene	13.115	105	65069	5.04	ug/l	85
86) p-Isopropyltoluene	13.233	119	36527	3.22	ug/l	97
89) n-Butylbenzene	13.557	91	39299	5.54	ug/l #	73
95) Naphthalene	15.070	128	144535	12.40	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066267.D
Acq On : 18 Mar 2021 19:38
Operator : JC/MD
Sample : M1687-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1052-1053

Quant Time: Mar 19 04:30:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

