

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066270.D
 Acq On : 18 Mar 2021 20:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 04:31:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.492	0.411	16.5	77	0.00
3 P	Chloromethane	0.870	0.672	22.8	80	0.00
4 C	Vinyl Chloride	0.710	0.613	13.7#	80	0.00
5 T	Bromomethane	0.406	0.367	9.6	88	0.00
6 T	Chloroethane	0.411	0.369	10.2	85	0.00
7 T	Trichlorofluoromethane	0.831	0.788	5.2	86	0.00
8 T	Diethyl Ether	0.340	0.311	8.5	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.440	0.426	3.2	89	0.00
10 T	Methyl Iodide	0.624	0.640	-2.6	94	0.00
11 T	Tert butyl alcohol	0.123	0.119	3.3	85	0.00
12 CM	1,1-Dichloroethene	0.469	0.430	8.3#	88	0.00
13 T	Acrolein	0.112	0.094	16.1	89	0.00
14 T	Allyl chloride	1.112	1.053	5.3	88	0.00
15 T	Acrylonitrile	0.368	0.335	9.0	80	0.00
16 T	Acetone	0.346	0.292	15.6	80	0.00
17 T	Carbon Disulfide	1.326	1.176	11.3	84	0.00
18 T	Methyl Acetate	0.968	0.990	-2.3	86	0.00
19 T	Methyl tert-butyl Ether	1.795	1.576	12.2	80	0.00
20 T	Methylene Chloride	0.541	0.525	3.0	91	0.00
21 T	trans-1,2-Dichloroethene	0.520	0.481	7.5	87	0.00
22 T	Diisopropyl ether	2.285	1.797	21.4	68	0.00
23 T	Vinyl Acetate	1.959	1.489	24.0	66	0.00
24 P	1,1-Dichloroethane	1.139	0.970	14.8	76	0.00
25 T	2-Butanone	0.538	0.434	19.3	69	0.00
26 T	2,2-Dichloropropane	0.941	0.688	26.9#	66	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.561	6.5	84	0.00
28 T	Bromochloromethane	0.602	0.484	19.6	73	0.00
29 T	Tetrahydrofuran	0.370	0.296	20.0	69	0.00
30 C	Chloroform	1.040	0.934	10.2#	81	0.00
31 T	Cyclohexane	1.079	0.773	28.4#	68	0.00
32 T	1,1,1-Trichloroethane	0.894	0.794	11.2	81	0.00
33 S	1,2-Dichloroethane-d4	0.720	0.568	21.1	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.319	0.322	-0.9	88	0.00
36 T	1,1-Dichloropropene	0.506	0.454	10.3	75	0.00
37 T	Ethyl Acetate	0.724	0.575	20.6	66	0.00
38 T	Carbon Tetrachloride	0.454	0.451	0.7	81	0.00
39 T	Methylcyclohexane	0.544	0.496	8.8	74	0.00
40 TM	Benzene	1.520	1.470	3.3	80	0.00
41 T	Methacrylonitrile	0.357	0.288	19.3	69	0.00
42 TM	1,2-Dichloroethane	0.552	0.498	9.8	74	0.00
43 T	Isopropyl Acetate	1.119	0.910	18.7	66	0.00
44 TM	Trichloroethene	0.369	0.373	-1.1	84	0.00
45 C	1,2-Dichloropropane	0.440	0.417	5.2#	77	0.00
46 T	Dibromomethane	0.245	0.251	-2.4	84	0.00
47 T	Bromodichloromethane	0.534	0.524	1.9	82	0.00
48 T	Methyl methacrylate	0.530	0.445	16.0	68	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
 Data File : VN066270.D
 Acq On : 18 Mar 2021 20:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 04:31:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 17 05:05:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	95	0.00
50 S	Toluene-d8	1.233	1.257	-1.9	89	0.00
51 T	4-Methyl-2-Pentanone	0.676	0.702	-3.8	81	0.00
52 CM	Toluene	0.881	0.949	-7.7#	88	0.00
53 T	t-1,3-Dichloropropene	0.604	0.594	1.7	79	0.00
54 T	cis-1,3-Dichloropropene	0.657	0.641	2.4	80	0.00
55 T	1,1,2-Trichloroethane	0.361	0.394	-9.1	90	0.00
56 T	Ethyl methacrylate	0.567	0.601	-6.0	83	0.00
57 T	1,3-Dichloropropane	0.651	0.687	-5.5	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.061	0.123	-101.6#	175#	0.00
59 T	2-Hexanone	0.482	0.519	-7.7	82	0.00
60 T	Dibromochloromethane	0.378	0.419	-10.8	91	0.00
61 T	1,2-Dibromoethane	0.359	0.400	-11.4	90	0.00
62 S	4-Bromofluorobenzene	0.445	0.471	-5.8	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.415	0.348	16.1	76	0.00
65 PM	Chlorobenzene	1.003	1.020	-1.7	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.384	-2.1	91	0.00
67 C	Ethyl Benzene	1.779	1.768	0.6#	86	0.00
68 T	m/p-Xylenes	0.660	0.687	-4.1	91	0.00
69 T	o-Xylene	0.634	0.659	-3.9	90	0.00
70 T	Styrene	1.052	1.172	-11.4	95	0.00
71 P	Bromoform	0.298	0.337	-13.1	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.504	3.435	2.0	89	0.00
74 T	N-amyl acetate	1.582	1.578	0.3	78	0.00
75 P	1,1,2,2-Tetrachloroethane	1.225	1.251	-2.1	93	0.00
76 T	1,2,3-Trichloropropane	1.291	1.169	9.5	84	0.00
77 T	Bromobenzene	0.919	0.914	0.5	95	0.00
78 T	n-propylbenzene	4.030	4.013	0.4	88	0.00
79 T	2-Chlorotoluene	2.461	2.423	1.5	89	0.00
80 T	1,3,5-Trimethylbenzene	2.907	2.912	-0.2	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.402	6.7	84	0.00
82 T	4-Chlorotoluene	2.438	2.516	-3.2	89	0.00
83 T	tert-Butylbenzene	2.535	2.414	4.8	90	0.00
84 T	1,2,4-Trimethylbenzene	2.842	2.867	-0.9	90	0.00
85 T	sec-Butylbenzene	3.251	3.247	0.1	89	0.00
86 T	p-Isopropyltoluene	2.851	2.874	-0.8	88	0.00
87 T	1,3-Dichlorobenzene	1.558	1.597	-2.5	95	0.00
88 T	1,4-Dichlorobenzene	1.631	1.599	2.0	92	0.00
89 T	n-Butylbenzene	2.538	2.336	8.0	81	0.00
90 T	Hexachloroethane	0.499	0.503	-0.8	98	0.00
91 T	1,2-Dichlorobenzene	1.534	1.540	-0.4	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.243	0.220	9.5	76	0.00
93 T	1,2,4-Trichlorobenzene	1.002	0.903	9.9	81	0.00
94 T	Hexachlorobutadiene	0.487	0.448	8.0	86	0.00
95 T	Naphthalene	2.932	2.580	12.0	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031821\
Data File : VN066270.D
Acq On : 18 Mar 2021 20:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 19 04:31:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031621W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 17 05:05:15 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.998	0.921	7.7	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6