

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
 Data File : VN071373.D
 Acq On : 18 Mar 2022 15:30
 Operator : JC\MD
 Sample : N1960-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-20220316

Manual Integrations
 APPROVED

Reviewed By :John Carlone 03/19/2022
 Supervised By :Mahesh Dadoda 03/21/2022

Quant Time: Mar 19 07:07:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.087	168	788423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1320976	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1243290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	445456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	505269	46.760	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.520%
35) Dibromofluoromethane	8.022	113	387937	46.539	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.080%
50) Toluene-d8	10.439	98	1623088	48.360	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.720%
62) 4-Bromofluorobenzene	12.727	95	588827	51.810	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.620%
Target Compounds						
16) Acetone	4.287	43	10109m	1.923	ug/l	
18) Methyl Acetate	4.869	43	9916	1.166	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
 Data File : VN071373.D
 Acq On : 18 Mar 2022 15:30
 Operator : JC\MD
 Sample : N1960-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-20220316

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/19/2022
 Supervised By : Mahesh Dadoda 03/21/2022

Quant Time: Mar 19 07:07:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

