

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
 Data File : VN071378.D
 Acq On : 18 Mar 2022 17:30
 Operator : JC\MD
 Sample : N1974-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B-3

Quant Time: Mar 19 07:09:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	775027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1304561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1241740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	461045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	502957	47.350	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.700%
35) Dibromofluoromethane	8.022	113	380203	46.185	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.360%
50) Toluene-d8	10.439	98	1606094	48.456	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.920%
62) 4-Bromofluorobenzene	12.727	95	606190	54.009	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.020%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	6621	1.281	ug/l #	82
18) Methyl Acetate	4.857	43	15013	1.541	ug/l	97
40) Benzene	8.457	78	202873	5.105	ug/l	100
52) Toluene	10.504	92	78895	3.224	ug/l	94
67) Ethyl Benzene	11.839	91	180957	3.838	ug/l	98
68) m/p-Xylenes	11.945	106	107318	6.222	ug/l	97
69) o-Xylene	12.275	106	46681	2.700	ug/l	92
73) Isopropylbenzene	12.575	105	12898	0.290	ug/l	97
80) 1,3,5-Trimethylbenzene	13.057	105	23479	0.679	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	74974	2.264	ug/l	98
95) Naphthalene	15.504	128	106585	9.599	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
Data File : VN071378.D
Acq On : 18 Mar 2022 17:30
Operator : JC\MD
Sample : N1974-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B-3

Quant Time: Mar 19 07:09:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

