

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
 Data File : VN071382.D
 Acq On : 18 Mar 2022 19:07
 Operator : JC\MD
 Sample : N2015-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-04

Quant Time: Mar 19 07:11:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	774246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1297485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1238691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	447446	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	498507	46.979	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.960%
35) Dibromofluoromethane	8.022	113	382016	46.658	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.320%
50) Toluene-d8	10.439	98	1626323	49.334	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.727	95	590755	52.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.840%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.440	62	58181	5.595	ug/l	90
16) Acetone	4.293	43	12586	2.438	ug/l	97
18) Methyl Acetate	4.851	43	11309	1.278	ug/l #	63
19) Methyl tert-butyl Ether	5.616	73	83611	2.869	ug/l #	78
27) cis-1,2-Dichloroethene	7.328	96	54573	5.416	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
Data File : VN071382.D
Acq On : 18 Mar 2022 19:07
Operator : JC\MD
Sample : N2015-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-04

Quant Time: Mar 19 07:11:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

