

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
 Data File : VN071385.D
 Acq On : 18 Mar 2022 20:20
 Operator : JC\MD
 Sample : N2015-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-07

Quant Time: Mar 19 07:12:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	744599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1262294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1185699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	407153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	484726	47.499	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.000%
35) Dibromofluoromethane	8.022	113	372468	46.760	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.520%
50) Toluene-d8	10.439	98	1575631	49.129	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.260%
62) 4-Bromofluorobenzene	12.727	95	557958	51.376	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.760%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.363	59	16607	8.284	ug/l #	88
16) Acetone	4.304	43	15784	3.179	ug/l	96
18) Methyl Acetate	4.851	43	10365	1.240	ug/l #	90
19) Methyl tert-butyl Ether	5.622	73	96065	3.428	ug/l	95
22) Diisopropyl ether	6.504	45	15207	0.475	ug/l #	84
27) cis-1,2-Dichloroethene	7.322	96	6906	0.713	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031822\
Data File : VN071385.D
Acq On : 18 Mar 2022 20:20
Operator : JC\MD
Sample : N2015-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-07

Quant Time: Mar 19 07:12:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

