

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031824\
 Data File : VN081452.D
 Acq On : 18 Mar 2024 16:03
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Mar 19 05:42:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 05:35:05 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 03/19/2024
 Supervised By : Mahesh Dadoda 03/19/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	460364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	758169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	702545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	344121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	595480	101.430	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 202.860%#			
35) Dibromofluoromethane	8.165	113	506989	105.958	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 211.920%#			
50) Toluene-d8	10.571	98	1889815	109.260	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 218.520%#			
62) 4-Bromofluorobenzene	12.853	95	703191	115.379	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 230.760%#			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.124	85	614271	99.095	ug/l	98
3) Chloromethane	2.359	50	567704	91.984	ug/l	98
4) Vinyl Chloride	2.512	62	693322	100.098	ug/l	98
5) Bromomethane	2.942	94	460781	94.067	ug/l	93
6) Chloroethane	3.112	64	467485	98.301	ug/l	97
7) Trichlorofluoromethane	3.495	101	1056411	102.337	ug/l	93
8) Diethyl Ether	3.959	74	359851	105.334	ug/l	77
9) 1,1,2-Trichlorotrifluo...	4.371	101	593408	101.170	ug/l #	87
10) Methyl Iodide	4.589	142	703427	120.428	ug/l	96
11) Tert butyl alcohol	5.518	59	484299	473.363	ug/l #	85
12) 1,1-Dichloroethene	4.342	96	553544	101.885	ug/l	86
13) Acrolein	4.177	56	376600	410.215	ug/l	100
14) Allyl chloride	5.018	41	699810	102.793	ug/l #	82
15) Acrylonitrile	5.712	53	1357843	496.095	ug/l	97
16) Acetone	4.418	43	841049	443.282	ug/l	93
17) Carbon Disulfide	4.712	76	1532805	98.313	ug/l	98
18) Methyl Acetate	5.018	43	602953	103.079	ug/l #	87
19) Methyl tert-butyl Ether	5.794	73	1835135	107.928	ug/l	96
20) Methylene Chloride	5.277	84	613648	98.736	ug/l #	84
21) trans-1,2-Dichloroethene	5.789	96	611003	100.863	ug/l	90
22) Diisopropyl ether	6.671	45	1740623	110.037	ug/l #	92
23) Vinyl Acetate	6.600	43	6951932	552.130	ug/l #	90
24) 1,1-Dichloroethane	6.571	63	1066769	100.687	ug/l	96
25) 2-Butanone	7.483	43	1625459	499.480	ug/l #	86
26) 2,2-Dichloropropane	7.488	77	1010523	109.968	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	710627	103.680	ug/l	85
28) Bromochloromethane	7.812	49	385733	89.386	ug/l #	55
29) Tetrahydrofuran	7.835	42	1129983	515.430	ug/l #	78
30) Chloroform	7.965	83	1165759	102.781	ug/l	97
31) Cyclohexane	8.259	56	923899	96.708	ug/l #	73
32) 1,1,1-Trichloroethane	8.171	97	1071463	104.124	ug/l #	93
36) 1,1-Dichloropropene	8.371	75	882930	109.650	ug/l	95
37) Ethyl Acetate	7.559	43	715269	103.735	ug/l	97
38) Carbon Tetrachloride	8.365	117	960643	108.905	ug/l	99
39) Methylcyclohexane	9.600	83	1074820	119.527	ug/l	85
40) Benzene	8.606	78	2553293	108.460	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031824\
 Data File : VN081452.D
 Acq On : 18 Mar 2024 16:03
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/19/2024
 Supervised By : Mahesh Dadoda 03/19/2024

Quant Time: Mar 19 05:42:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 05:35:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	380450	104.334	ug/l #	91
42) 1,2-Dichloroethane	8.671	62	880663	106.299	ug/l	98
43) Isopropyl Acetate	8.688	43	1230389	107.651	ug/l #	91
44) Trichloroethene	9.353	130	712040	107.541	ug/l	94
45) 1,2-Dichloropropane	9.624	63	630414	110.495	ug/l	97
46) Dibromomethane	9.712	93	464047	107.102	ug/l #	87
47) Bromodichloromethane	9.888	83	938206	110.752	ug/l #	98
48) Methyl methacrylate	9.682	41	590366	111.987	ug/l #	85
49) 1,4-Dioxane	9.694	88	258597	2041.213	ug/l #	81
51) 4-Methyl-2-Pentanone	10.447	43	3835898	550.013	ug/l	91
52) Toluene	10.629	92	1698621	113.523	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	1051847	114.989	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	1105784	112.749	ug/l	93
55) 1,1,2-Trichloroethane	11.018	97	638772	106.079	ug/l	93
56) Ethyl methacrylate	10.876	69	993179	99.988	ug/l #	84
57) 1,3-Dichloropropane	11.165	76	1091349	110.556	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	1017930	268.656	ug/l #	86
59) 2-Hexanone	11.194	43	2911114	564.574	ug/l	92
60) Dibromochloromethane	11.359	129	741019	111.896	ug/l	99
61) 1,2-Dibromoethane	11.470	107	678862	110.578	ug/l	97
64) Tetrachloroethene	11.106	164	716955	106.762	ug/l	92
65) Chlorobenzene	11.894	112	1790633	107.816	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.965	131	687317	111.660	ug/l	96
67) Ethyl Benzene	11.965	91	3360632	116.488	ug/l	98
68) m/p-Xylenes	12.070	106	2593691	239.818	ug/l	94
69) o-Xylene	12.400	106	1247481	117.492	ug/l	96
70) Styrene	12.412	104	2096232	122.414	ug/l	96
71) Bromoform	12.582	173	521708	112.789	ug/l #	96
73) Isopropylbenzene	12.700	105	3332020	113.592	ug/l	98
74) N-amyl acetate	12.494	43	1221772	115.095	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.941	83	841944	97.623	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	827844m	100.799	ug/l	
77) Bromobenzene	12.982	156	754297	104.097	ug/l	72
78) n-propylbenzene	13.035	91	4010676	115.778	ug/l	96
79) 2-Chlorotoluene	13.129	91	2267320	109.358	ug/l	93
80) 1,3,5-Trimethylbenzene	13.176	105	2796554	115.427	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	339872	111.576	ug/l	86
82) 4-Chlorotoluene	13.223	91	2295424	112.296	ug/l	94
83) tert-Butylbenzene	13.441	119	2337071	111.862	ug/l	95
84) 1,2,4-Trimethylbenzene	13.482	105	2823356	115.946	ug/l	97
85) sec-Butylbenzene	13.617	105	3486991	116.752	ug/l	95
86) p-Isopropyltoluene	13.729	119	2920609	119.627	ug/l	96
87) 1,3-Dichlorobenzene	13.735	146	1417510	106.473	ug/l	96
88) 1,4-Dichlorobenzene	13.812	146	1416330	102.742	ug/l	96
89) n-Butylbenzene	14.059	91	2644571	122.157	ug/l	97
90) Hexachloroethane	14.335	117	514728	111.324	ug/l	89
91) 1,2-Dichlorobenzene	14.106	146	1328929	103.710	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.723	75	189998	103.517	ug/l	71
93) 1,2,4-Trichlorobenzene	15.394	180	851569	114.935	ug/l	98
94) Hexachlorobutadiene	15.506	225	352486	108.429	ug/l	94
95) Naphthalene	15.641	128	2817582	117.360	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	835235	112.073	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031824\
Data File : VN081452.D
Acq On : 18 Mar 2024 16:03
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024

Quant Time: Mar 19 05:42:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 05:35:05 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031824\
Data File : VN081452.D
Acq On : 18 Mar 2024 16:03
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024


Quant Time: Mar 19 05:42:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 05:35:05 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:168 Resp: 460364
Ion Ratio Lower Upper
168 100
99 62.2 59.9 89.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#2
Dichlorodifluoromethane
Concen: 99.095 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion: 85 Resp: 614271
Ion Ratio Lower Upper
85 100
87 31.8 15.3 45.9





#3
Chloromethane
Concen: 91.984 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 567704
Ion Ratio Lower Upper
50 100
52 32.1 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#4
Vinyl Chloride
Concen: 100.098 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 62 Resp: 693322
Ion Ratio Lower Upper
62 100
64 29.2 22.7 34.1









#5
Bromomethane
Concen: 94.067 ug/l
RT: 2.942 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 94 Resp: 46078
Ion Ratio Lower Upper
94 100
96 93.3 69.2 103.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#6
Chloroethane
Concen: 98.301 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 64 Resp: 467485
Ion Ratio Lower Upper
64 100
66 31.2 26.4 39.6









#7
Trichlorofluoromethane
Concen: 102.337 ug/l
RT: 3.495 min Scan# 263
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:101 Resp: 105641
Ion Ratio Lower Upper
101 100
103 62.7 54.4 81.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#8
Diethyl Ether
Concen: 105.334 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 74 Resp: 359851
Ion Ratio Lower Upper
74 100
45 82.8 53.3 159.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 101.170 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:101 Resp: 593408
Ion Ratio Lower Upper
101 100
85 44.7 38.1 57.1
151 76.5 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#10
Methyl Iodide
Concen: 120.428 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion:142 Resp: 703427
Ion Ratio Lower Upper
142 100
127 49.2 37.0 55.4
141 14.7 12.5 18.7





#11
Tert butyl alcohol
Concen: 473.363 ug/l
RT: 5.518 min Scan# 606
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 59 Resp: 484299
Ion Ratio Lower Upper
59 100
57 4.9 8.6 12.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024




#12
1,1-Dichloroethene
Concen: 101.885 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Tgt Ion: 96 Resp: 553544
Ion Ratio Lower Upper
96 100
61 148.5 138.2 207.4
98 63.8 49.2 73.8





#13
Acrolein
Concen: 410.215 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 376600
Ion Ratio Lower Upper
56 100
55 69.0 55.3 82.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#14
Allyl chloride
Concen: 102.793 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion: 41 Resp: 699810
Ion Ratio Lower Upper
41 100
39 84.9 57.7 86.5
76 47.8 26.8 40.2#





#15
Acrylonitrile
Concen: 496.095 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 53 Resp: 135784
Ion Ratio Lower Upper
53 100
52 82.8 64.8 97.2
51 36.5 26.9 40.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#16
Acetone
Concen: 443.282 ug/l
RT: 4.418 min Scan# 419
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 43 Resp: 841049
Ion Ratio Lower Upper
43 100
58 34.6 24.8 37.2









#17
Carbon Disulfide
Concen: 98.313 ug/l
RT: 4.712 min Scan# 409
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 76 Resp: 153280
Ion Ratio Lower Upper
76 100
78 8.7 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#18
Methyl Acetate
Concen: 103.079 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 43 Resp: 602953
Ion Ratio Lower Upper
43 100
74 27.7 17.3 25.9





#19

Methyl tert-butyl Ether

Concen: 107.928 ug/l

RT: 5.794 min Scan# 653

Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 73 Resp: 183513

Ion Ratio Lower Upper

73 100





57 20.9 18.2 27.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/19/2024

Supervised By :Mahesh Dadoda 03/19/2024



#20

Methylene Chloride

Concen: 98.736 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Tgt Ion: 84 Resp: 613648




Ion Ratio Lower Upper


84 100

49 107.0 108.1 162.1#

51 34.1 32.9 49.3

86 62.6 48.6 73.0









#21
trans-1,2-Dichloroethene
Concen: 100.863 ug/l
RT: 5.789 min Scan# 611
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 96 Resp: 61100
Ion Ratio Lower Upper
96 100
61 124.2 112.8 169.2
98 62.6 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#22
Diisopropyl ether
Concen: 110.037 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion: 45 Resp: 1740623
Ion Ratio Lower Upper
45 100
43 54.5 41.8 62.8
87 33.4 18.1 27.1#
59 12.9 9.0 13.6





#23
 Vinyl Acetate
 Concen: 552.130 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100




Tgt Ion: 43 Resp: 695193
 Ion Ratio Lower Upper
 43 100
 86 12.6 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :John Carlone 03/19/2024
 Supervised By :Mahesh Dadoda 03/19/2024




#24
 1,1-Dichloroethane
 Concen: 100.687 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.006 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03



Tgt Ion: 63 Resp: 1066769
 Ion Ratio Lower Upper
 63 100
 98 7.4 2.9 8.6
 100 4.4 1.9 5.7









#25
2-Butanone
Concen: 499.480 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 43 Resp: 1625459
Ion Ratio Lower Upper
43 100
72 31.3 19.5 29.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#26
2,2-Dichloropropane
Concen: 109.968 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 77 Resp: 1010523
Ion Ratio Lower Upper
77 100
97 22.7 11.5 34.5









#27
 cis-1,2-Dichloroethene
 Concen: 103.680 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100

Tgt Ion: 96 Resp: 71062
 Ion Ratio Lower Upper
 96 100
 61 135.6 0.0 325.4
 98 65.5 0.0 130.4




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/19/2024
 Supervised By :Mahesh Dadoda 03/19/2024



#28
 Bromochloromethane
 Concen: 89.386 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03

Tgt Ion: 49 Resp: 385733
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 1.6#
 130 85.3 42.3 63.5#





#29

Tetrahydrofuran

Concen: 515.430 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 112998

Ion Ratio Lower Upper

42 100

72 51.9 31.1 46.7





71 49.1 28.3 42.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/19/2024

Supervised By :Mahesh Dadoda 03/19/2024



#30

Chloroform

Concen: 102.781 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN081452.D




Acq: 18 Mar 2024 16:03


Tgt Ion: 83 Resp: 1165759

Ion Ratio Lower Upper

83 100


85 66.4 51.0 76.6





#31
Cyclohexane
Concen: 96.708 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion: 56 Resp: 92389
Ion Ratio Lower Upper
56 100
69 32.0 20.1 30.1
84 97.9 56.9 85.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#32
1,1,1-Trichloroethane
Concen: 104.124 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 97 Resp: 1071463
Ion Ratio Lower Upper
97 100
99 66.9 53.4 80.0
61 43.3 43.8 65.8





#33

1,2-Dichloroethane-d4

Concen: 101.430 ug/l

RT: 8.582 min Scan# 1126

Delta R.T. 0.006 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 65 Resp: 595480

Ion Ratio Lower Upper

65 100





67 52.1 0.0 102.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/19/2024

Supervised By :Mahesh Dadoda 03/19/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.006 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03




Tgt Ion:114 Resp: 758169


Ion Ratio Lower Upper

114 100

63 19.2 0.0 48.0

88 14.7 0.0 34.8





#35

Dibromofluoromethane

Concen: 105.958 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:	113	Resp:	506989
Ion	Ratio	Lower	Upper
113	100		
111	104.2	82.2	123.4
192	17.8	12.5	18.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/19/2024

Supervised By :Mahesh Dadoda 03/19/2024



#36

1,1-Dichloropropene

Concen: 109.650 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03


Tgt Ion:	75	Resp:	882930
Ion	Ratio	Lower	Upper
75	100		
110	36.4	16.0	48.0
77	31.3	25.8	38.6





#37
Ethyl Acetate
Concen: 103.735 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 715269
Ion Ratio Lower Upper
43 100
61 14.3 10.6 15.8
70 11.3 8.0 12.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#38
Carbon Tetrachloride
Concen: 108.905 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion:117 Resp: 960643
Ion Ratio Lower Upper
117 100
119 93.3 73.4 110.2
121 30.7 24.4 36.6





#39
Methylcyclohexane
Concen: 119.527 ug/l
RT: 9.600 min Scan# 11300
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 83 Resp: 1074820
Ion Ratio Lower Upper
83 100
55 67.7 65.7 98.5
98 49.7 33.8 50.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024




#40
Benzene
Concen: 108.460 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Tgt Ion: 78 Resp: 2553293
Ion Ratio Lower Upper
78 100
77 24.2 20.4 30.6





#41
Methacrylonitrile
Concen: 104.334 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 41 Resp: 380450
Ion Ratio Lower Upper
41 100
39 60.2 48.4 72.6
67 85.5 56.8 85.2
52 34.6 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#42
1,2-Dichloroethane
Concen: 106.299 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion: 62 Resp: 880663
Ion Ratio Lower Upper
62 100
98 9.9 0.0 18.0





#43
Isopropyl Acetate
Concen: 107.651 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 1230389
Ion Ratio Lower Upper
43 100
61 28.6 19.9 29.9
87 15.8 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#44
Trichloroethene
Concen: 107.541 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:130 Resp: 712040
Ion Ratio Lower Upper
130 100
95 99.2 0.0 210.4





#45

1,2-Dichloropropane

Concen: 110.495 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 630414

Ion Ratio Lower Upper

63 100





65 30.9 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/19/2024

Supervised By :Mahesh Dadoda 03/19/2024



#46

Dibromomethane

Concen: 107.102 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03




Tgt Ion: 93 Resp: 464047


Ion Ratio Lower Upper

93 100

95 82.6 68.1 102.1


174 93.1 57.2 85.8#





#47
Bromodichloromethane
Concen: 110.752 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion: 83 Resp: 938200
Ion Ratio Lower Upper
83 100
85 64.1 52.3 78.5
127 8.3 4.6 7.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024




#48
Methyl methacrylate
Concen: 111.987 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Tgt Ion: 41 Resp: 590366
Ion Ratio Lower Upper
41 100
69 96.3 62.2 93.2
39 65.4 47.6 71.4





#49
1,4-Dioxane
Concen: 2041.213 ug/l
RT: 9.694 min Scan# 1316
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 88 Resp: 258597
Ion Ratio Lower Upper
88 100
43 25.8 30.1 45.1
58 63.6 63.9 95.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#50
Toluene-d8
Concen: 109.260 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 98 Resp: 1889815
Ion Ratio Lower Upper
98 100
100 65.2 51.4 77.0









#51
4-Methyl-2-Pentanone
Concen: 550.013 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 3835898
Ion Ratio Lower Upper
43 100
58 41.9 29.4 44.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#52
Toluene
Concen: 113.523 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 92 Resp: 1698621
Ion Ratio Lower Upper
92 100
91 173.3 139.4 209.0









#53
t-1,3-Dichloropropene
Concen: 114.989 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 105184
Ion Ratio Lower Upper
75 100
77 31.8 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#54
cis-1,3-Dichloropropene
Concen: 112.749 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion: 75 Resp: 1105784
Ion Ratio Lower Upper
75 100
77 31.3 23.3 34.9
39 50.0 44.6 66.8





#55
1,1,2-Trichloroethane
Concen: 106.079 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:	97	Resp:	638772
Ion	Ratio	Lower	Upper
97	100		
83	82.7	72.9	109.3
85	56.0	50.2	75.2
99	63.7	49.6	74.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#56
Ethyl methacrylate
Concen: 99.988 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:	69	Resp:	993179
Ion	Ratio	Lower	Upper
69	100		
41	58.3	61.0	91.6#
39	35.3	31.8	47.6





#57
1,3-Dichloropropane
Concen: 110.556 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 76 Resp: 1091349
Ion Ratio Lower Upper
76 100
78 32.3 25.5 38.3

Manual Integrations

APPROVED




Reviewed By :John Carlone 03/19/2024


Supervised By :Mahesh Dadoda 03/19/2024



#58
2-Chloroethyl Vinyl ether
Concen: 268.656 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 63 Resp: 1017930
Ion Ratio Lower Upper
63 100
106 30.3 18.6 28.0#









#59
2-Hexanone
Concen: 564.574 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 2911114
Ion Ratio Lower Upper
43 100
58 57.3 25.9 77.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#60
Dibromochloromethane
Concen: 111.896 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion:129 Resp: 741019
Ion Ratio Lower Upper
129 100
127 77.6 39.4 118.2





#61
1,2-Dibromoethane
Concen: 110.578 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion:107 Resp: 678862
Ion Ratio Lower Upper
107 100
109 94.4 73.4 110.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024




#62
4-Bromofluorobenzene
Concen: 115.379 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03



Tgt Ion: 95 Resp: 703191
Ion Ratio Lower Upper
95 100
174 77.6 0.0 120.4
176 74.6 0.0 121.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.870 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:117 Resp: 70254

Ion Ratio Lower Upper

117 100

82 53.2 52.7 79.1





119 32.3 25.3 37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/19/2024

Supervised By :Mahesh Dadoda 03/19/2024



#64

Tetrachloroethene

Concen: 106.762 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.000 min

Lab File: VN081452.D

Acq: 18 Mar 2024 16:03

Tgt Ion:164 Resp: 716955




Ion Ratio Lower Upper


164 100

166 127.8 105.8 158.6

129 100.3 89.6 134.4

131 93.8 84.9 127.3









#65
Chlorobenzene
Concen: 107.816 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:112 Resp: 179063
Ion Ratio Lower Upper
112 100
114 32.5 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 111.660 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:131 Resp: 687317
Ion Ratio Lower Upper
131 100
133 95.3 46.8 140.3
119 64.6 34.6 103.9









#67
Ethyl Benzene
Concen: 116.488 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 91 Resp: 336063
Ion Ratio Lower Upper
91 100
106 31.0 23.8 35.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#68
m/p-Xylenes
Concen: 239.818 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:106 Resp: 2593691
Ion Ratio Lower Upper
106 100
91 202.4 169.0 253.6









#69
o-Xylene
Concen: 117.492 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:106 Resp: 124748
Ion Ratio Lower Upper
106 100
91 217.2 112.1 336.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#70
Styrene
Concen: 122.414 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion:104 Resp: 2096232
Ion Ratio Lower Upper
104 100
78 49.8 43.8 65.8
103 55.6 44.2 66.2





#71
Bromoform
Concen: 112.789 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100




Tgt Ion:173 Resp: 521708
Ion Ratio Lower Upper
173 100
175 48.9 23.1 69.2
254 0.1 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03



Tgt Ion:152 Resp: 344121
Ion Ratio Lower Upper
152 100
115 60.5 32.3 96.8
150 196.1 0.0 339.8









#73
Isopropylbenzene
Concen: 113.592 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:105 Resp: 3332020
Ion Ratio Lower Upper
105 100
120 26.1 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#74
N-ethyl acetate
Concen: 115.095 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 43 Resp: 1221772
Ion Ratio Lower Upper
43 100
70 46.0 30.1 45.1#
55 24.9 17.0 25.4
61 25.4 17.5 26.3









#75
1,1,2,2-Tetrachloroethane
Concen: 97.623 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 83 Resp: 841944
Ion Ratio Lower Upper
83 100
131 10.3 4.1 12.3
85 64.6 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#76
1,2,3-Trichloropropane
Concen: 100.799 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 75 Resp: 827844
Ion Ratio Lower Upper
75 100
77 189.9 91.0 272.8









#77
Bromobenzene
Concen: 104.097 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:156 Resp: 754297
Ion Ratio Lower Upper
156 100
77 208.4 138.1 414.3
158 97.1 46.5 139.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#78
n-propylbenzene
Concen: 115.778 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion: 91 Resp: 4010676
Ion Ratio Lower Upper
91 100
120 22.4 10.3 30.9









#79
2-Chlorotoluene
Concen: 109.358 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 2267320
Ion Ratio Lower Upper
91 100
126 33.3 14.8 44.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#80
1,3,5-Trimethylbenzene
Concen: 115.427 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:105 Resp: 2796554
Ion Ratio Lower Upper
105 100
120 48.2 23.0 69.0









#81
trans-1,4-Dichloro-2-butene
Concen: 111.576 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.006 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument : MSVOA_N
ClientSampleId : VSTDICC100

Tgt Ion:	75	Resp:	339872
Ion Ratio	Lower	Upper	
75	100		
53	105.1	71.4	107.2
89	45.7	33.4	50.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#82
4-Chlorotoluene
Concen: 112.296 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:	91	Resp:	2295424
Ion Ratio	Lower	Upper	
91	100		
126	33.9	15.3	45.9









#83
tert-Butylbenzene
Concen: 111.862 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:119 Resp: 233707
Ion Ratio Lower Upper
119 100
91 68.9 36.8 110.4
134 26.6 12.4 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#84
1,2,4-Trimethylbenzene
Concen: 115.946 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:105 Resp: 2823356
Ion Ratio Lower Upper
105 100
120 43.7 20.9 62.7









#85
 sec-Butylbenzene
 Concen: 116.752 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Tgt Ion:105 Resp: 348699
 Ion Ratio Lower Upper
 105 100
 134 19.5 8.6 25.8




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/19/2024
 Supervised By :Mahesh Dadoda 03/19/2024



#86
 p-Isopropyltoluene
 Concen: 119.627 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03

Tgt Ion:119 Resp: 2920609
 Ion Ratio Lower Upper
 119 100
 134 26.3 12.4 37.2
 91 24.2 13.5 40.4









#87
1,3-Dichlorobenzene
Concen: 106.473 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:146 Resp: 1417510
Ion Ratio Lower Upper
146 100
111 42.2 23.9 71.7
148 64.3 32.6 97.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#88
1,4-Dichlorobenzene
Concen: 102.742 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:146 Resp: 1416330
Ion Ratio Lower Upper
146 100
111 40.0 22.6 67.7
148 62.9 31.9 95.5









#89
n-Butylbenzene
Concen: 122.157 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion: 91 Resp: 264457
Ion Ratio Lower Upper
91 100
92 52.3 26.7 80.1
134 25.0 10.8 32.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#90
Hexachloroethane
Concen: 111.324 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:117 Resp: 514728
Ion Ratio Lower Upper
117 100
201 68.0 30.0 90.0





#91
1,2-Dichlorobenzene
Concen: 103.710 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:146 Resp: 1328929
Ion Ratio Lower Upper
146 100
111 42.9 23.9 71.8
148 63.9 31.4 94.0

Manual Integrations

APPROVED




Reviewed By :John Carlone 03/19/2024


Supervised By :Mahesh Dadoda 03/19/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 103.517 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03


Tgt Ion: 75 Resp: 189998
Ion Ratio Lower Upper
75 100
155 85.6 31.8 95.4
157 108.5 41.1 123.4





#93
 1,2,4-Trichlorobenzene
 Concen: 114.935 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03




Instrument : MSVOA_N
 ClientSampleId : VSTDIC100



Tgt Ion:180 Resp: 851569
 Ion Ratio Lower Upper
 180 100
 182 94.2 46.6 139.7
 145 32.0 16.8 50.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/19/2024
 Supervised By :Mahesh Dadoda 03/19/2024



#94
 Hexachlorobutadiene
 Concen: 108.429 ug/l
 RT: 15.506 min Scan# 2304
 Delta R.T. 0.000 min
 Lab File: VN081452.D
 Acq: 18 Mar 2024 16:03


Tgt Ion:225 Resp: 352486
 Ion Ratio Lower Upper
 225 100
 223 62.5 33.3 99.8
 227 61.0 33.1 99.2





#95
Naphthalene
Concen: 117.360 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100



Tgt Ion:128 Resp: 281758
Ion Ratio Lower Upper
128 100
127 13.0 10.0 15.0
129 10.9 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/19/2024
Supervised By :Mahesh Dadoda 03/19/2024



#96
1,2,3-Trichlorobenzene
Concen: 112.073 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.000 min
Lab File: VN081452.D
Acq: 18 Mar 2024 16:03

Tgt Ion:180 Resp: 835235
Ion Ratio Lower Upper
180 100
182 96.8 46.5 139.3
145 33.6 17.8 53.4

